




CHARGE SENSITIVE PREAMPLIFIER/ SHAPING AMPLIFIER

A-203



VOLTAGE AMPLIFIER/ LOW LEVEL DISCRIMINATOR






Models A-203 and A-206 are a Charge Sensitive Preamplifier/Shaping Amplifier and a matching Voltage Amplifier/Low Level Discriminator developed especially for instrumentation employing solid state detectors, proportional counters, photomultipliers or any charge producing detectors in the pulse height analysis or pulse counting mode of operation.

While these units were specifically designed for satellite instrumentation, the following unique characteristics make them equally useful for space, laboratory and commercial applications:

- Small Size (16 Pin Dual In-Line Package) allows mounting close to the detector.
- · Power required is typically 18 milliwatts.
- · Single power supply voltage.
- Low noise.

- · Pole-zero cancellation (external).
- · Unipolar and bipolar outputs.
- Both Pulse Height Analysis and Pulse Counting mode of operation.
- High Reliability.
- · One year warranty.

THE A-203 / A-206 COMPLETE SYSTEM



AMPTEK HIGH RELIABILITY SCREENING

1. PRECAP VISUAL:

MIL-STD-883, method 2017. Low Magnification, High Magnification.

2. SEALING:

Welded, hermetic seal.

Date code and serial number.

3. STAMPING: 4. STABILIZATION BAKE:

MIL-STD-883, method 1008, Condition C. +150 °C, 24 hours minimum. MIL-STD-883, method 1010, Condition C min. T = -65 °C to +150 °C.

5. TEMPERATURE CYCLE:

10 minutes each extreme, 5 minutes maximum transfer time.

6. CENTRIFUGE:

MIL-STD-883, method 2001, Condition B. Yl Axis; 10,000 G's. As per specifications.

7. ELECTRICAL TEST: 8. BURN-IN TEST:

MIL-STD-883, method 1015, 160 hours at + 125°C.

9. FINE LEAK TEST: 10. GROSS LEAK TEST: MIL-STD-883, method 1014, Condition A. Rejection if leak rate in excess of 5×10^{-7} cc/sec.

MIL-STD-883, method 1014, Condition C. Fluoro Carbon: Rejection if stream of

11. ELECTRICAL TEST:

bubbles is present.
As per specifications.

12. EXTERNAL VISUAL:

MIL-STD-883, method 2009.

AMPTEK INC.

6 DE ANGELO DRIVE, BEDFORD, MA 01730 U.S.A. TEL: (617) 275-2242 FAX: (617) 275-3470

■ 0807411 0000292 T96 ■

CHARGE SENSITIVE PREAMPLIFIER / SHAPING AMPLIFIER

SPECIFICATIONS

 $(V_S = + 15V, T = 25 °C)$

INPUT CHARACTERISTICS

SENSITIVITY: 1) For positive unipolar output (Pin 8): 210 mv/Mev (Si); 262 mv/Mev (Ge); 4.8x1012v/coulomb; 0.76 μv/electron

> 2) For bipolar output (Pin 9): 115 mv/Mev (Si); 144 mv/Mev (Ge); 2.6x10¹²v/coulomb; 0.42 μv/electron.

NOISE: 1) Unipolar: (Nominal)

8 Kev FWHM (Si); 6.4 Kev FWHM (Ge); 1.4x10⁻¹⁶ coulomb rms; 875

electrons rms

2) Bipolar:

12 Kev FWHM (Si); 10 Kev FWHM (Ge); 2.2x10-18 coulomb rms

NOISE SLOPE: (Typical)

100 ev/pf (Si); 80 ev/pf (Ge); 1.81x10-18 coulombs/pf

DYNAMIC INPUT

CAPACITANCE: 3,500 pf

POLARITY: Negative PROTECTION:

Back-to-back diodes to ground

OUTPUT CHARACTERISTICS

1) CHARGE SENSITIVE PREAMPLIFIER OUTPUT-Pin 13:

Rise time: 50ns Fall time: 30μs **AC Output**

impedance: 50Ω 2) SHAPING AMPLIFIER

Time constants: 250ns

Positive Unipolar.

Pin 8:

AC output impedance: 85Ω Dynamic range: 75% of Vs

Pole-zero cancellation: External 1MΩ

between Pins 12 and 13

Bipolar, Pin 9:

AC output impedance: 2kΩ Dynamic range: 35% of Vs

GENERAL

OPERATING VOLTAGE: + 10 to + 18 VDC OPERATING CURRENT: 1.4 ma Quiescent @15V TEMPERATURE: -55° to +70°C operational PACKAGE: 16 Pin Dual In-Line (Metal) SCREENING: AMPTEK HIGH RELIABILITY

WARRANTY: One year

OPERATING NOTES

The A-203 has two parts:

1) A Charge Sensitive Preamplifier (CSP).

2) A Shaping Amplifier (SA).


The output of the CSP is internally connected to the input of the SA.

- Power to the CSP and SA are provided separately in order to provide maximum flexibility for independent operation. However, for normal operation Pin 11 and Pin 14 should be connected to Vs. Supply is internally bypassed. Care should be taken in circuit layout and in some applications power supply decoupling may be helpful. If the leads of the A-203 are left long, or the unit is socketed, the unconnected Pins 2, 3, 6, 7, 10, 15, 16 should be grounded in order to minimizé pick-up.
- The sensitivity of the CSP is defined by: G = 44/C in mv/Mev (Si), where C is the feedback capacitor (in pf). The feedback capacitor in the A-203 is 2pf. If a lower sensitivity

is desired, an external capacitor may be added between Pins 1 and 13. A capacitor with good temperature stability should be used and the leads should be kept short.

- The detector must be capacitively coupled to Pin 1 with a capacitor of adequate voltage rating.
- If the output of the CSP (Pin 13) is to be connected to external circuitry it must be coupled with an external capacitor. This output can drive several feet of unterminated coaxial cable.
- The SA has two outputs: 1) Unipolar (Pin 8) which is single integration - single differentiation. 2) Bipolar (Pin 9) which is single integration - double differentiation. Either output can drive the A-206, which has an internal input coupling capacitor. If the unipolar output (Pin 8) is to be connected to other external circuitry it MUST be capacitively coupled.
- If pole-zero cancellation is required at the unipolar output in order to minimize the baseline shift, a $1M\Omega$ resistor must be externally added between Pins 12 and 13.
- The A-203 can be tested with a pulser by using a small capacitor to inject a negative test charge into the input. The unit will respond to the negative-going edge of either a square wave or a tail pulse with long fall time (> 10 μ s). In either case the negative-going transition should be less than 20 ns. Charge transfer in the circuit is according to Q=CV, where Q=total amount of charge delivered to input, C = test capacitor and V = amplitude of the pulse. Use only a small capacitor in this circuit (1-10pf). DO NOT connect a low impedance pulser through a large capacitor when testing as this will overdrive the input and may damage the unit.


Example: A 22 mv tail pulse across a 2pf test capacitor is equivalent to 1 Mev energy loss in silicon (3.5 ev/pair) and will produce a 210 mv unipolar pulse at Pin 8.



Fall time $> 10\mu$ s, or square wave Amplitude: 22mv = 1Mev (Si)

 $Q = CV = (2x10^{-12}f) \times (22x10^{-3}v) = 4.4x10^{-14}coulomb = 1Mev(Si)$

A-203 LOGIC AND PIN CONFIGURATION



Charge Sensitive Preamplifier Input Pin 1

Pins 2, 3, 6, 7, 10, 15, 16 No Connection

Pins 4, 5 Ground and case

Pin 8 **Unipolar Shaping Amplifier Output** Pin 9 Bipolar Shaping Amplifier Output

Pin 11 Vs Shaping Amplifier (+ 10V to + 18V)

Pin 12 Shaping Amplifier Input

Charge Sensitive Preamplifier Output **Pin 13** Pin 14 Vs C.S. Preamplifier (+ 10V to + 18V)

■ 22P EPS0000 1147080

VOLTAGE AMPLIFIER / LOW LEVEL DISCRIMINATOR

SPECIFICATIONS

 $(V_s = + 15V, T = 25$ °C)

1) VOLTAGE AMPLIFIER

Gain: x 10 nominal, non inverting Input polarity: 1) Positive Unipolar; Dynamic

range: 85% of Vs

Bipolar, If a bipolar input is used, an external resistor $(R_B = 330K\Omega)$ must be connected between Pin 3 and ground. Dynamic range: 65% of Vs

AC Output Impedance: 20Ω

Integral Nonlinearity

at full output: ≤ 0.5%

2) DISCRIMINATOR

Input Polarity: Positive or Bipolar

Output pulse: Positive; 5µs wide; 90% of Vs. The

output pulse width can be increased by the addition of an external capacitor (1 to 10 pf) between Pins

6 and 8.

Rise Time: 20 ns

Reference voltage: Pin 11, VR; Range: +4.5 to +5.5 VDC (Impedance >2 M Ω)

Variable voltage: Pin 10, V_v ; $(V_v < V_R)$

Discrimination Level: $V_R - V_V$ (Impedance >2M Ω)

GENERAL

OPERATING VOLTAGE: + 10 to + 18VDC

OPERATING CURRENT: 0.8 ma Quiescent @15V


TEMPERATURE: -55°C to +70°C operational

PACKAGE: 16 Pin Dual In-Line (Metal)

SCREENING: AMPTEK HIGH RELIABILITY

WARRANTY: One year.

A-206 LOGIC AND PIN CONFIGURATION



Pin 1 Ground and Case

Pins 2, 7, 14, 15 No Connection

Bipolar Input Bias Pin 3

 V_s Amplifier (+ 10V to + 18V) Pin 4

Vs Discriminator (+ 10V to + 18V) Pin 5

Discriminator Pulse Width Adjust Pin 6

Discriminator Pulse Width Adjust Pin 8

Pin 9 **Discriminator Output**

Pin 10 + V_V variable (V_V < V_R)

Pin 11 + VR reference (+ 4.5 to + 5.5 VDC)

Discriminator Input Pin 12


Amplifier Output Pin 13

Pin 16 **Amplifier Input**

OPERATING NOTES


- The A-206 has two parts:
 - 1) A matching Voltage Amplifier (VA) to the A-203 or A-225
 - 2) A Lower Level Discriminator (LLD)

- The output of the VA is NOT internally connected to the LLD. Pin 13 MUST be connected to Pin 12 if the LLD is to be used.
- Power to the VA and LLD is provided separately in order to provide maximum flexibility for independent operation. However, for normal operation Pins 4 and 5 should be connected to Vs.
- The VA has an internal input coupling capacitor and no other external capacitor is needed when connecting to either unipolar or bipolar outputs of the A-203 or the A-225.
- If the output of the VA (Pin 13) is to be connected to external circuitry it MUST be coupled with an external capacitor. This output can drive several feet of unterminated coaxial cable.
- For bipolar input pulses to the VA connect a $330k\Omega$ resistor from Pin 3 to ground. This biases the amplifier for maximum dynamic range.
- The LLD requires a reference voltage VR and a variable voltage V_V. Set V_R at a level between + 4.5 and + 5.5 volts and V_v to a lower voltage than V_B. The difference, V_B-V_v, is the discrimination level at which the unit will trigger. For example, if $V_R = +5V$ and $V_V = +3.5v$ any pulse of amplitude greater than + 1.5v will trigger the discriminator.
- The LLD output pulse is typically a 5 µsec wide positive pulse. A wider pulse can be obtained by connecting a capacitor (1 to 10 pf) between Pins 6 and 8. Leave these pins unconnected if the standard pulse width is acceptable.
- A typical circuit to provide bias to the LLD is shown below.



A temperature stable voltage reference should be used.


TEST BOARD (ACTUAL SIZE)




PC-236 test board for the A-203/A206.

0807411 0000294 869


THE A-203/A-206 COMPLETE SYSTEM AND TYPICAL WAVEFORMS




TYPICAL APPLICATIONS



Typical connection of a photomultiplier to the A-203



Typical connection of a solid state detector to the A-203



AMPTEK INC.

6 DE ANGELO DRIVE, BEDFORD, MA 01730 U.S.A. TEL: (617) 275-2242 FAX: (617) 275-3470

■ 0807411 0000295 7T5 **■**